

Rely+On™ PeraSafe™

LANXESS

High Level Disinfectant for use with Medical Devices
For use on invasive and non-invasive medical devices.

Applications and instructions for use

These instructions for use (IFU) should be read in conjunction with the product label and SDS.

Rely+On™ PeraSafe™ is for use on rigid and flexible endoscopes, catheters, transducers and other thermolabile instruments and equipment. For a list of manufacturer approvals, contact your local representative.

Not to be used on materials constructed of copper, brass, mild steel, natural rubber, nitrile rubber, natural dental bristle brushes or EPDM rubber.

Always wear suitable eye protection when handling powder and in-use solutions of Rely+On™ PeraSafe™. Where contact with dusts of the powder are foreseen, ensure adequate ventilation and/or respiratory protection (e.g. surgical mask).

Rely+On™ PeraSafe™ solutions can be used for up to 20 immersions or 12 hours, whichever is first.

Instruments should be thoroughly pre-cleaned with a medical grade pH-neutral detergent to remove all visible soiling, followed by a thorough rinse with water to remove all detergent prior to disinfection with Rely+On™ PeraSafe™. Ensure all surfaces of the equipment are in contact with the solution.

Heavily soiled instruments may give faster degradation of the peracetic acid level present, this may result in a reduction of the maximum number of immersions.

Composition: c. 0.25% w/v at 16.2 g/litre peracetic acid generated in-situ from tetraacetylenediamine (TAED) and sodium percarbonate; surfactants; corrosion inhibitor; excipients.



Rely+On™ PeraSafe™ should be stored in the sealed original container, preferably at temperatures not exceeding 25°C and a 60% relative humidity. Storage for short periods at temperatures not exceeding 30° C and 65% relative humidity will not affect product performance.

Prepare only at the labelled concentration of -16.2 g/litre. Do NOT make up over-strength solutions which are subsequently diluted.

Rely+On™ PeraSafe™

LANXESS

High Level Disinfectant for use with Medical Devices
For use on invasive and non-invasive medical devices.

Applications and instructions for use

These instructions for use (IFU) should be read in conjunction with the product label and SDS.

Rely+On™ PeraSafe™ is for use on rigid and flexible endoscopes, catheters, transducers and other thermolabile instruments and equipment. For a list of manufacturer approvals, contact your local representative.

Not to be used on materials constructed of copper, brass, mild steel, natural rubber, nitrile rubber, natural dental bristle brushes or EPDM rubber.

Always wear suitable eye protection when handling powder and in-use solutions of Rely+On™ PeraSafe™. Where contact with dusts of the powder are foreseen, ensure adequate ventilation and/or respiratory protection (e.g. surgical mask).

Rely+On™ PeraSafe™ solutions can be used for up to 20 immersions or 12 hours, whichever is first.

Instruments should be thoroughly pre-cleaned with a medical grade pH-neutral detergent to remove all visible soiling, followed by a thorough rinse with water to remove all detergent prior to disinfection with Rely+On™ PeraSafe™. Ensure all surfaces of the equipment are in contact with the solution.

Heavily soiled instruments may give faster degradation of the peracetic acid level present, this may result in a reduction of the maximum number of immersions.

Composition: c. 0.25% w/v at 16.2 g/litre peracetic acid generated in-situ from tetraacetylenediamine (TAED) and sodium percarbonate; surfactants; corrosion inhibitor; excipients.



Rely+On™ PeraSafe™ should be stored in the sealed original container, preferably at temperatures not exceeding 25°C and a 60% relative humidity. Storage for short periods at temperatures not exceeding 30° C and 65% relative humidity will not affect product performance.

Prepare only at the labelled concentration of -16.2 g/litre. Do NOT make up over-strength solutions which are subsequently diluted.

Rely+On™ PeraSafe™

LANXESS

High Level Disinfectant for use with Medical Devices
For use on invasive and non-invasive medical devices.

Applications and instructions for use

These instructions for use (IFU) should be read in conjunction with the product label and SDS.

Rely+On™ PeraSafe™ is for use on rigid and flexible endoscopes, catheters, transducers and other thermolabile instruments and equipment. For a list of manufacturer approvals, contact your local representative.

Not to be used on materials constructed of copper, brass, mild steel, natural rubber, nitrile rubber, natural dental bristle brushes or EPDM rubber.

Always wear suitable eye protection when handling powder and in-use solutions of Rely+On™ PeraSafe™. Where contact with dusts of the powder are foreseen, ensure adequate ventilation and/or respiratory protection (e.g. surgical mask).

Rely+On™ PeraSafe™ solutions can be used for up to 20 immersions or 12 hours, whichever is first.

Instruments should be thoroughly pre-cleaned with a medical grade pH-neutral detergent to remove all visible soiling, followed by a thorough rinse with water to remove all detergent prior to disinfection with Rely+On™ PeraSafe™. Ensure all surfaces of the equipment are in contact with the solution.

Heavily soiled instruments may give faster degradation of the peracetic acid level present, this may result in a reduction of the maximum number of immersions.

Composition: c. 0.25% w/v at 16.2 g/litre peracetic acid generated in-situ from tetraacetylenediamine (TAED) and sodium percarbonate; surfactants; corrosion inhibitor; excipients.



Rely+On™ PeraSafe™ should be stored in the sealed original container, preferably at temperatures not exceeding 25°C and a 60% relative humidity. Storage for short periods at temperatures not exceeding 30° C and 65% relative humidity will not affect product performance.

Prepare only at the labelled concentration of -16.2 g/litre. Do NOT make up over-strength solutions which are subsequently diluted.

Rely+On™ PeraSafe™

LANXESS

High Level Disinfectant for use with Medical Devices
For use on invasive and non-invasive medical devices.

Applications and instructions for use

These instructions for use (IFU) should be read in conjunction with the product label and SDS.

Rely+On™ PeraSafe™ is for use on rigid and flexible endoscopes, catheters, transducers and other thermolabile instruments and equipment. For a list of manufacturer approvals, contact your local representative.

Not to be used on materials constructed of copper, brass, mild steel, natural rubber, nitrile rubber, natural dental bristle brushes or EPDM rubber.

Always wear suitable eye protection when handling powder and in-use solutions of Rely+On™ PeraSafe™. Where contact with dusts of the powder are foreseen, ensure adequate ventilation and/or respiratory protection (e.g. surgical mask).

Rely+On™ PeraSafe™ solutions can be used for up to 20 immersions or 12 hours, whichever is first.

Instruments should be thoroughly pre-cleaned with a medical grade pH-neutral detergent to remove all visible soiling, followed by a thorough rinse with water to remove all detergent prior to disinfection with Rely+On™ PeraSafe™. Ensure all surfaces of the equipment are in contact with the solution.

Heavily soiled instruments may give faster degradation of the peracetic acid level present, this may result in a reduction of the maximum number of immersions.

Composition: c. 0.25% w/v at 16.2 g/litre peracetic acid generated in-situ from tetraacetylenediamine (TAED) and sodium percarbonate; surfactants; corrosion inhibitor; excipients.



Rely+On™ PeraSafe™ should be stored in the sealed original container, preferably at temperatures not exceeding 25°C and a 60% relative humidity. Storage for short periods at temperatures not exceeding 30° C and 65% relative humidity will not affect product performance.

Prepare only at the labelled concentration of -16.2 g/litre. Do NOT make up over-strength solutions which are subsequently diluted.

Preparation and use of activated Rely+On™ PeraSafe™ solution for instrument disinfection

The performance and safety of Rely+On™ PeraSafe™ has been validated under the conditions of use described within this document. Always follow the instructions to ensure correct operation of the product.

Rely+On™ PeraSafe™ is effective against bacteria, bacterial spores, mycobacteria, viruses and yeasts, with a contact time of 10 minutes.

Preparation and use of Rely+On™ PeraSafe™ solution (1.62% w/v)

1. Weigh using an appropriate balance 16.2 g of Rely+On™ PeraSafe™ powder per 1 litre of lukewarm water (30-35°C). Do not prepare solutions of less than 1 Litre.
2. Activation takes 15 minutes. Stir periodically throughout this activation period. Solution becomes colourless after around 30 minutes. A small amount of undissolved material is expected.
3. Rely+On™ PeraSafe™ solutions should be used and stored at room temperature (20 ± 5°C). However, Rely+On™ PeraSafe™ can be used immediately after activation as the solution cools to room temperature.
4. Rely+On™ PeraSafe™ solution may be used in a static bath. Immerse instruments for exactly 10 minutes, after which they must be thoroughly rinsed in sterile water, being sure to remove all traces of the Rely+On™ PeraSafe™ solution, then dried.

It is recommended that Rely+On™ PeraSafe™ solutions are not prepared more than one hour prior to usage. Used Rely+On™ PeraSafe™ solution may be discarded, without special precautions, in the normal liquid waste disposal system in accordance with local authority regulations.



Antec International Limited
Windham Road, Chilton Industrial Estate, Sudbury, Suffolk,
CO10 2XD, United Kingdom
Tel: +44 (0) 1787 377 305 Fax: +44 (0) 1787 310 846
relyondisinfection@lanxess.com www.relyondisinfection.com

Marketed by: LANXESS India Private Limited
Business Unit: Material Protection Products (MPP)
Registered Office: LANXESS House, Plot no: A 162-164,
Road No 27, MIDC, Wagle Estate, Thane (W) - 400 604,
Maharashtra, India
Phone: +91 22 2587 1000 (Board)
Fax: +91 22 25871287 www.lanxess.in

Imported by: Vishal Surgical Equipment Co. Pvt. Ltd.
23 Wallers Road, Melaram Complex,
Chennai -600 002, India
Tel: 044-28549809
Email: customercare@vishalsurgical.co.in
Emergency phone no.: +91-2645-224140

Preparation and use of activated Rely+On™ PeraSafe™ solution for instrument disinfection

The performance and safety of Rely+On™ PeraSafe™ has been validated under the conditions of use described within this document. Always follow the instructions to ensure correct operation of the product.

Rely+On™ PeraSafe™ is effective against bacteria, bacterial spores, mycobacteria, viruses and yeasts, with a contact time of 10 minutes.

Preparation and use of Rely+On™ PeraSafe™ solution (1.62% w/v)

1. Weigh using an appropriate balance 16.2 g of Rely+On™ PeraSafe™ powder per 1 litre of lukewarm water (30-35°C). Do not prepare solutions of less than 1 Litre.
2. Activation takes 15 minutes. Stir periodically throughout this activation period. Solution becomes colourless after around 30 minutes. A small amount of undissolved material is expected.
3. Rely+On™ PeraSafe™ solutions should be used and stored at room temperature (20 ± 5°C). However, Rely+On™ PeraSafe™ can be used immediately after activation as the solution cools to room temperature.
4. Rely+On™ PeraSafe™ solution may be used in a static bath. Immerse instruments for exactly 10 minutes, after which they must be thoroughly rinsed in sterile water, being sure to remove all traces of the Rely+On™ PeraSafe™ solution, then dried.

It is recommended that Rely+On™ PeraSafe™ solutions are not prepared more than one hour prior to usage. Used Rely+On™ PeraSafe™ solution may be discarded, without special precautions, in the normal liquid waste disposal system in accordance with local authority regulations.



Antec International Limited
Windham Road, Chilton Industrial Estate, Sudbury, Suffolk,
CO10 2XD, United Kingdom
Tel: +44 (0) 1787 377 305 Fax: +44 (0) 1787 310 846
relyondisinfection@lanxess.com www.relyondisinfection.com

Marketed by: LANXESS India Private Limited
Business Unit: Material Protection Products (MPP)
Registered Office: LANXESS House, Plot no: A 162-164,
Road No 27, MIDC, Wagle Estate, Thane (W) - 400 604,
Maharashtra, India
Phone: +91 22 2587 1000 (Board)
Fax: +91 22 25871287 www.lanxess.in

Imported by: Vishal Surgical Equipment Co. Pvt. Ltd.
23 Wallers Road, Melaram Complex,
Chennai -600 002, India
Tel: 044-28549809
Email: customercare@vishalsurgical.co.in
Emergency phone no.: +91-2645-224140

©2021 LANXESS. Rely+On™ PeraSafe™, LANXESS and the LANXESS Logo are trademarks of LANXESS Deutschland GmbH or its affiliates
All trademarks are registered in many countries worldwide.

62267593/Psafe/IFU/India/14.10.21/Z

©2021 LANXESS. Rely+On™ PeraSafe™, LANXESS and the LANXESS Logo are trademarks of LANXESS Deutschland GmbH or its affiliates
All trademarks are registered in many countries worldwide.

62267593/Psafe/IFU/India/14.10.21/Z

Preparation and use of activated Rely+On™ PeraSafe™ solution for instrument disinfection

The performance and safety of Rely+On™ PeraSafe™ has been validated under the conditions of use described within this document. Always follow the instructions to ensure correct operation of the product.

Rely+On™ PeraSafe™ is effective against bacteria, bacterial spores, mycobacteria, viruses and yeasts, with a contact time of 10 minutes.

Preparation and use of Rely+On™ PeraSafe™ solution (1.62% w/v)

1. Weigh using an appropriate balance 16.2 g of Rely+On™ PeraSafe™ powder per 1 litre of lukewarm water (30-35°C). Do not prepare solutions of less than 1 Litre.
2. Activation takes 15 minutes. Stir periodically throughout this activation period. Solution becomes colourless after around 30 minutes. A small amount of undissolved material is expected.
3. Rely+On™ PeraSafe™ solutions should be used and stored at room temperature (20 ± 5°C). However, Rely+On™ PeraSafe™ can be used immediately after activation as the solution cools to room temperature.
4. Rely+On™ PeraSafe™ solution may be used in a static bath. Immerse instruments for exactly 10 minutes, after which they must be thoroughly rinsed in sterile water, being sure to remove all traces of the Rely+On™ PeraSafe™ solution, then dried.

It is recommended that Rely+On™ PeraSafe™ solutions are not prepared more than one hour prior to usage. Used Rely+On™ PeraSafe™ solution may be discarded, without special precautions, in the normal liquid waste disposal system in accordance with local authority regulations.



Antec International Limited
Windham Road, Chilton Industrial Estate, Sudbury, Suffolk,
CO10 2XD, United Kingdom
Tel: +44 (0) 1787 377 305 Fax: +44 (0) 1787 310 846
relyondisinfection@lanxess.com www.relyondisinfection.com

Marketed by: LANXESS India Private Limited
Business Unit: Material Protection Products (MPP)
Registered Office: LANXESS House, Plot no: A 162-164,
Road No 27, MIDC, Wagle Estate, Thane (W) - 400 604,
Maharashtra, India
Phone: +91 22 2587 1000 (Board)
Fax: +91 22 25871287 www.lanxess.in

Imported by: Vishal Surgical Equipment Co. Pvt. Ltd.
23 Wallers Road, Melaram Complex,
Chennai -600 002, India
Tel: 044-28549809
Email: customercare@vishalsurgical.co.in
Emergency phone no.: +91-2645-224140

Preparation and use of activated Rely+On™ PeraSafe™ solution for instrument disinfection

The performance and safety of Rely+On™ PeraSafe™ has been validated under the conditions of use described within this document. Always follow the instructions to ensure correct operation of the product.

Rely+On™ PeraSafe™ is effective against bacteria, bacterial spores, mycobacteria, viruses and yeasts, with a contact time of 10 minutes.

Preparation and use of Rely+On™ PeraSafe™ solution (1.62% w/v)

1. Weigh using an appropriate balance 16.2 g of Rely+On™ PeraSafe™ powder per 1 litre of lukewarm water (30-35°C). Do not prepare solutions of less than 1 Litre.
2. Activation takes 15 minutes. Stir periodically throughout this activation period. Solution becomes colourless after around 30 minutes. A small amount of undissolved material is expected.
3. Rely+On™ PeraSafe™ solutions should be used and stored at room temperature (20 ± 5°C). However, Rely+On™ PeraSafe™ can be used immediately after activation as the solution cools to room temperature.
4. Rely+On™ PeraSafe™ solution may be used in a static bath. Immerse instruments for exactly 10 minutes, after which they must be thoroughly rinsed in sterile water, being sure to remove all traces of the Rely+On™ PeraSafe™ solution, then dried.

It is recommended that Rely+On™ PeraSafe™ solutions are not prepared more than one hour prior to usage. Used Rely+On™ PeraSafe™ solution may be discarded, without special precautions, in the normal liquid waste disposal system in accordance with local authority regulations.



Antec International Limited
Windham Road, Chilton Industrial Estate, Sudbury, Suffolk,
CO10 2XD, United Kingdom
Tel: +44 (0) 1787 377 305 Fax: +44 (0) 1787 310 846
relyondisinfection@lanxess.com www.relyondisinfection.com

Marketed by: LANXESS India Private Limited
Business Unit: Material Protection Products (MPP)
Registered Office: LANXESS House, Plot no: A 162-164,
Road No 27, MIDC, Wagle Estate, Thane (W) - 400 604,
Maharashtra, India
Phone: +91 22 2587 1000 (Board)
Fax: +91 22 25871287 www.lanxess.in

Imported by: Vishal Surgical Equipment Co. Pvt. Ltd.
23 Wallers Road, Melaram Complex,
Chennai -600 002, India
Tel: 044-28549809
Email: customercare@vishalsurgical.co.in
Emergency phone no.: +91-2645-224140

©2021 LANXESS. Rely+On™ PeraSafe™, LANXESS and the LANXESS Logo are trademarks of LANXESS Deutschland GmbH or its affiliates
All trademarks are registered in many countries worldwide.

62267593/Psafe/IFU/India/14.10.21/Z

©2021 LANXESS. Rely+On™ PeraSafe™, LANXESS and the LANXESS Logo are trademarks of LANXESS Deutschland GmbH or its affiliates
All trademarks are registered in many countries worldwide.

62267593/Psafe/IFU/India/14.10.21/Z